

HANBY TPH TEST KITS- DETECTION LIMITS LIST

Petroleum Substance	SOIL (PPM)	WATER (PPM)	
Unleaded Gasoline	1	0.1	
Diesel fuel	1	0.1	
Crude oil	1	0.1	
Heating oil	1	0.1	
PCBs (<i>polychlorinated biphenyl</i>)	1	0.01	(10PPB)
Regular Gasoline	1	0.1	
Benzene	1	0.02	
Toluene	1	0.02	(20PPB)
Xylene	1	0.02	
Jet Fuel	1	0.1	
Super Unleaded Gasoline	1	0.1	
Kerosene	1	0.1	
Oil / Grease in Water	5	1	
Mineral Oil	20	0.25	
Piceance Condensate	1	0.1	
Deepwater Horizon Crude Oil	1	0.1	
Fuel Oil	1	0.1	
Motor Oil	20	0.01	
Waste Oil	10	0.01	
Transformer Oil	20	0.1	
Lubricating Oil	20	0.1	
DDT (Insecticide)	1	0.01	
Arochlor 1248	1	0.01	
Arochlor 1016	1	0.01	
PAH Compounds (N, Ac, An, Ph, Fl)	1	0.01	
<i>(Naphalene,Acenaphthene,Anthracene, Phenanthrene & Flouranthene)</i>			
Napthalene	1	0.01	
JP 4	1	0.01	
JP 5	1	0.01	
Valdez Crude Oil	1	0.01	